

# Densah® Bur Tracking Log



	Pilot	VT1525	VT1828	VS2228	VT2535	VT2838	VS3238	VT3545	VT3848	VS4248	VT4555	VT4858	VS5258
1													
2													
3													
4													
5													
6													
7													
8													
9													
10													
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													

## Densah® Burs Maintenance & Replacement

**Surgical drills and burs should be replaced after **12-20 osteotomies**\* or sooner when they are dulled or worn.**

**Cutting and Densifying must be done under constant water irrigation. A pumping motion is required to prevent overheating. It is recommended to keep a spare set of Densah® Burs on hand in the event replacement is needed during a surgery.**

\*Chacon GE, Bower DL, Larson PR, et al. Heat production by three implant drill systems after repeated drilling and sterilization. J Oral Maxillofac Surg. 2006;64(2):265-269.