



Use large block display to compare Bur identification system when using the schematic below for proper Bur usage

● VT5 Set ○ VT8 Set ● VS8 Set

| Densifying Mode CCW (800-1500) RPMs / Cutting Mode CW (800-1500) RPMs | | | | | | | | | | | | | | | | | |
|---|---------|---------|---------------------------|--------------|---------------|---------------|---------------|---------------------------|--|--------------|--------------|---------------|---------------|---------------|---------------|---------------|---------------------------|
| Hiossen | | | ETII, ETIII, HSII Fixture | | | | | | | | | | | | | | |
| | | | Soft Bone | | | | | | Hard Bone (Mandible) | | | | | | | | |
| | | | | | | | | | In densifying mode make sure your osteotomy is 1.0 mm deeper than the actual implant final length. In extreme hard bone, utilize DAC (Densify After Cut) Protocol. Click here to view PDF. | | | | | | | | |
| Geometry | Major Ø | Minor Ø | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Densah® Bur Block Display | Pilot | Bur 1 | Bur 2 | Bur 3 | Bur 4 | Bur 5 | Bur 6 | Bur 7 | Densah® Bur Block Display |
| Slow Taper | 3.5 | | Pilot | VT1525 (2.0) | VT2535* (3.0) | — | — | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT2838* (3.3) | — | — | — | — | |
| Slow Taper | 4.0 | | Pilot | VT1828 (2.3) | VT2838* (3.3) | — | — | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VS3238* (3.5) | — | — | — | |
| Slow Taper | 4.5 | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT3545* (4.0) | — | | Pilot | VT1525 (2.0) | VT2535 (3.0) | VT2838 (3.3) | VT3545 (4.0) | VT3848* (4.3) | — | — | |
| Slow Taper | 5.0 | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848* (4.3) | — | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VS4248* (4.5) | — | |
| Slow Taper | 6.0 | | Pilot | VT1828 (2.3) | VT2838 (3.3) | VT3848 (4.3) | VT4858* (5.3) | | Pilot | VT1828 (2.3) | VT2535 (3.0) | VT2838 (3.3) | VT3545 (4.0) | VT3848 (4.3) | VT4858 (5.3) | VS5258* (5.5) | |

*Denotes implant placement.

NOTE: Surgeon preference overrides this suggestive protocol